

# Resource Summary Report

Generated by [RRID](#) on Sep 19, 2026

## BioSpherix: C Shuttle Glove Box hood

RRID:SCR\_021082

Type: Tool

---

### Proper Citation

BioSpherix: C Shuttle Glove Box hood (RRID:SCR\_021082)

---

### Resource Information

**URL:** <https://biospherix.com/c-shuttle/>

**Proper Citation:** BioSpherix: C Shuttle Glove Box hood (RRID:SCR\_021082)

**Description:** Oxygen controlled hood for handling cells under physiologic or pathophysiologic conditions. Double door entry allows import/export of materials without perturbation of experimental conditions. Optional control of carbon dioxide as well. Receives C-Chamber Incubator Subchamber to transfer cells without exposure to room air.

**Abbreviations:** C-Shuttle

**Synonyms:** BioSpherix C-Shuttle Glove Box, Hypoxic Glove Box, C-Shuttle laboratory glove box

**Resource Type:** instrument resource

**Keywords:** BioSpherix, Physioxia, Physioxic, Hypoxia, Hypoxic, Glove box, oxygen, cell culture hood, carbon dioxide, instrument, equipment, USEDit

**Availability:** Commercially available

**Specification URL:** [https://raw.githubusercontent.com/SciCrunch/RRID-Instruments/refs/heads/main/PDF/SCR\\_021082.pdf](https://raw.githubusercontent.com/SciCrunch/RRID-Instruments/refs/heads/main/PDF/SCR_021082.pdf)

**Resource Name:** BioSpherix: C Shuttle Glove Box hood

**Resource ID:** SCR\_021082

**Record Creation Time:** 20220129T080353+0000

**Record Last Update:** 20260912T195950+0000

---

### Ratings and Alerts

No rating or validation information has been found for BioSpherix: C Shuttle Glove Box hood.

No alerts have been found for BioSpherix: C Shuttle Glove Box hood.

---

## Data and Source Information

**Source:** [SciCrunch Registry](#)

---

## Usage and Citation Metrics

We have not found any literature mentions for this resource.